

Summary Specification	
BIL. SEBUTHARGA/QUOTATION RE	<u>UBD/Q/196/2024 (FOS) - [T]</u>
TARIKH DIKLANAKAN :	<u>15 Oktober 2024 [Selasa]</u>
TARIKH TUTUP :	<u>29 Oktober 2024 [Selasa]</u> JAM: 2 Petang
DOCUMENT FEE :	B\$ 10.00

IMPORTANT

The Official PDF quotation/tender form will be send via email once UBD Official receipt issued and send to vendors/supplier for confirmation of payment received.

Mode of Payments > REFER IN UBD WEBSITE

Please take note that the document fee for each of the New and Extended quotations will be BND 10.00 each

Important Note: Please ensure that you provide the correct details when making the payment.

Please email proof of the above payment to helpdesk.epay@ubd.edu.bn , finance@ubd.edu.bn, eprocurement.support@ubd.edu.bn.

TITLE :		<u>TO SUPPLY, DELIVER, AND INSTALL BENCH-TOP SHAKING INCUBATOR, FACULTY OF SCIENCE, UNIVERSITI BRUNEI DARUSSALAM</u>
No.	Quantity	Descriptions
1	1 unit	<p>Bench-Top Shaking Incubator</p> <p><u>Features:</u></p> <ul style="list-style-type: none"> • The constant temperature incubator shaker is a small bench-top shaking incubator that combines the functions of gas bath constant temperature and horizontal oscillation • A variety of optional trays meet the placement and fixation of various flasks, beakers, petri dishes, test tubes and other containers • Its ingenious combination of precise temperature control and low frequency and high amplitude oscillation functions; and it has characteristics of small size, convenient operation, simple maintenance and so on <p><u>Technical specifications:</u></p> <ul style="list-style-type: none"> • CAPACITY: 160 L • CIRCULATION: FORCED CONVECTION • SHAKE MODE: CONVOLUTION • ORBIT DIAMETER: 20 MM • CYCLOTRON FREQUENCY RANGE: 50 – 250 RPM/MIN • TEMPERATURE RANGE: RT+5 ~ 60 0C • TEMPERATURE FLUCTUATION: < +0.5 0C (37 0C) • TEMPERATURE RESOLUTION: 0.1 0C • TEMPERATURE UNIFORMITY: < +0.5 0C (37 0C) • EXTERNAL DIMENSION: 632(W) X 502(D) X 512(H) MM • CONTROLLER: MICROCOMPUTER • DISPLAY: LARGE HIGH DEFINITION LCD • TIMER 1MIN~99H59MIN/∞ • SAFETY DEVICE: RUNNING SAFELY: HEATING AND SHAKING WILL BE AUTOMATICALLY STOPPED ONCE THE LID IS OPEN. • VOLTAGE/FREQUENCY: AC 220/110V, 50/60HZ • N.W/G.W.: 41/58