| Summary Specification | | | |
|---|---|---|---|
| BIL. SEBUTHARGA/QUOTATION REF.: | | UBD/Q/300/2024 (FIT) - [T] | |
| TARIKH DIIKLANKAN : | 18 Februari 2025 [Selasa] | | |
| TARIKH TUTUP : | 04 Mac 2025 [Selasa] | JAM: 2 Petang | |
| | | DOCUMENT FEE : | B$ 10.00 |

**IMPORTANT**

**The Official PDF quotation/tender form will be send via email once UBD Official receipt issued and send to vendors/supplier for confirmation of payment received.**

**Mode of Payments > REFER IN UBD WEBSITE**

**Please take note that the document fee for each of the New and Extended quotations will be BND 10.00 each**

**Important Note: Please ensure that you provide the correct details when making the payment.**

Please email proof of the above payment to helpdesk.epay@ubd.edu.bn. , finance@ubd.edu.bn,eprocurement.support@ubd.edu.bn.

| TITLE : | | TO SUPPLY AND DELIVER TUBE FURNACES TO FACULTY OF INTEGRATED (FIT), UNIVERSITI BRUNEI DARUSSALAM FOR RESEARCH GRANT TITLE; INTERMEDIATE TEMPERATURE SOLID OXIDE ELECTROLYZER CELLS (SOECS) FOR GREEN HYDROGEN FOR ENERGY SUSTAINABILITY |
|---|---|---|
| **No.** | **Quantity** | **Descriptions** |
| 1 | 1 unit | **Spit type tube Furnace**<br>Maximum Temperature: 1200 deg C,<br>Tube size: 60 x 1000 mm,<br>Heating zone length: 300 mm.<br>Inclusive: Installation of 20A power circuit breaker for power supply |
| 2 | 1 unit | **Tube Furnace,**<br>Maximum Temperature: 1700 deg C,<br>Tube size: 80 x 1200 mm,<br>Heating zone length: 440 mm,<br>Heating element: MoSi2,<br>Power: 3 kW<br>Including accessories: Vacuum pump (1),<br>float flowmeter (1),<br>Flanges seal (1).<br>Installation of 20A power circuit breaker for power supply. |